



VIBROCONTROL 1100

Safety and reliability related values according to DIN EN ISO 13849-1

Safety and reliability related values MTTF, PL and Category according to DIN EN ISO 13849-1 have been evaluated for VC-1100 with the following results:

Device: VC-1100 C01 and C11

Parameter	Value (40°C)	Value (50°C)
MTTF	640.148 h ~ 73 years	414.950 h ~ 47 years
PL	c	c
Category	1	1

Device: VC-1100 C02 and C12

Parameter	Value (40°C)	Value (50°C)
MTTF	696.209 h ~ 79 years	541.741 h ~ 62 years
PL	c	c
Category	1	1

Device: VC-1100 C01/CCS and C11/CCS

Parameter	Value (40°C)	Value (50°C)
MTTF	638.674 h ~ 73 years	414.143 h ~ 47 years
PL	c	c
Category	1	1

Device: VC-1100 C02/CCS and C12/CCS

Parameter	Value (40°C)	Value (50°C)
MTTF	694.466 h ~ 79 years	540.784 h ~ 62 years
PL	c	c
Category	1	1

Test conditions:

- Operating temperatures: 40°C und 50°C.
- Environmental conditions: Ground Benign, Controlled



Note:

The results of this assessment are valid when the following procedures are followed:

- The Relays of the VC-1100 have to be operated in the „normally energised” mode.
- The analog 4-20 mA signal industry-standard current loops must be used.
- The OK Relay has to be used as system function.
- The VC-1100 system must be protected against erroneous change in configuration

Abbreviations:

MTTF	Mean Time To Failure
PL	Performance Level From PL "a" (highest failure probability) to PL "e" (lowest failure probability)
Category	Category (CAT) Classification of the safety related parts of a control system in respect of their resistance to faults and their subsequent behaviour in the fault condition, and which is achieved by the structural arrangement of the parts, fault detection and/or by their reliability
Ground Benign, Controlled	Nearly zero environmental stress with optimum engineering operation and maintenance.

Note:

The harmonized standard EN 954-1 is still valid (end of 2011), but it is considered technically outdated. It was replaced by the current harmonized standards EN ISO 13849-1. More detailed information about safety and reliability values can be obtained on request from Brüel & Kjær Vibro (info@bkvibro.com).

Internet/Ha En/Safety & Reliability VC-1100

Brüel & Kjær Vibro A/S
2850 Nærum – Denmark
Tel.: +45 7741 2500
Fax: +45 45 80 2937
E-mail: info@bkvibro.com

Brüel & Kjær Vibro GmbH
64293 Darmstadt – Germany
Tel.: +49 (0) 6151 428 1100
Fax: +49 (0) 6151 428 1200
E-mail: info@bkvibro.com